SCINTILLATION CRYSTAL

CAUTION: AVOID THERMAL AND MECHANICAL SHOCK - THIS CRYSTAL WILL PROBABLY CRACK IF SHOCK IS SEVERE.

This crystal unit is designed to be used with end-window photomultipliers such as the RCA type 5819, DuMont 6292, or EMI 6262, etc.

Silicone fluid or a similar material should be used between surfaces to provide good optical coupling.

This Crystal unit, 2" ϕ x 2"

Serial No.

Type Caesium Iodide (Thallium Activated)

~~has been inspected and tested for operation as a radiation detector.~~

22-3-61

NUCLEAR ENTERPRISES (G.B.) LTD.
Bankhead Medway, Sighthill,
Edinburgh, 11.